

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004735  
12:28 21/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:22  
STERILIZING 3:12  
136.0 C 220kPa 3:12  
136.4 C 224kPa 3:43  
136.3 C 222kPa 4:13  
136.2 C 220kPa 4:43  
136.4 C 223kPa 5:13  
136.2 C 219kPa 5:43  
136.2 C 225kPa 6:13  
136.2 C 223kPa 6:43  
136.2 C 223kPa 6:43  
VENTING 6:43  
AIR DRYING 7:32  
CYCLE COMPLETE 8:54

-----  
Digital Signature #  
C14E5AB2C1814354  
-----