
Midmark M11 - v1.0.5
Total Cycles: 4277

Name: RITTER2
Sterilizer ID: V2313228
01/22/2025 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 71.5 F 0.1
4:00 95.1 F 0.6
6:00 102.2 F 2.6
8:00 140.6 F 5.5
10:00 166.3 F 9.2
12:00 182.6 F 13.3
14:00 207.7 F 4.4
16:00 219.0 F 7.8
18:00 229.5 F 11.6
20:00 228.6 F 7.2
22:00 240.9 F 12.2
24:00 239.0 F 10.2
26:00 251.4 F 16.3
28:00 262.2 F 22.6
30:00 272.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.4 F 29.0
2:00 272.7 F 29.6
3:00 271.8 F 29.0
4:00 272.2 F 29.0

Min 271.6 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 184.8 F 0.0
10:00 193.9 F 0.0
15:00 199.2 F 0.0
20:00 201.8 F 0.0
25:00 206.1 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 30:24
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

01/22/2025 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____