
Midmark M11 - v1.0.0
Total Cycles: 2316

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2022 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.1 F 0.0
2:00 149.4 F 1.7
4:00 172.1 F 4.6
6:00 188.5 F 8.9
8:00 208.1 F 9.1
10:00 221.9 F 6.8
12:00 233.0 F 11.4
14:00 233.5 F 7.4
16:00 242.3 F 13.1
18:00 240.9 F 11.2
20:00 253.3 F 17.6
22:00 264.7 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.0 F 28.9
2:00 272.3 F 29.4
3:00 271.9 F 29.3
4:00 271.7 F 28.8

Min 271.5 F 28.7
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 196.9 F 0.0
10:00 205.3 F 0.0
15:00 211.8 F 0.0
20:00 215.1 F 0.0
25:00 220.2 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:51
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

09/13/2022 09:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____