
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002374
10:32 13/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:00
STERILIZING 3:02
136.0 C 221kPa 3:02
136.4 C 226kPa 3:32
136.1 C 225kPa 4:02
136.4 C 226kPa 4:32
136.3 C 225kPa 5:02
136.4 C 225kPa 5:32
136.3 C 222kPa 6:02
136.4 C 226kPa 6:32
136.4 C 226kPa 6:32
VENTING 6:32
AIR DRYING 7:27
CYCLE COMPLETE 37:27

Digital Signature #
5C7C94F149B428A1
