
Midmark M11 - v1.0.5
Total Cycles: 700

Name: RITTER2
Sterilizer ID: V2313228
01/17/2022 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.3 F 0.0
2:00 101.6 F 0.8
4:00 122.5 F 2.6
6:00 151.1 F 5.5
8:00 181.5 F 9.1
10:00 192.8 F 13.3
12:00 208.9 F 6.3
14:00 222.3 F 10.8
16:00 227.6 F 7.2
18:00 239.2 F 12.6
20:00 238.9 F 11.0
22:00 251.6 F 17.2
24:00 262.7 F 23.8
26:00 273.0 F 30.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 271.8 F 30.2
2:00 272.7 F 30.5
3:00 272.5 F 30.2
4:00 272.1 F 29.8

Min 271.2 F 29.4
Max 272.8 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.9 F 0.0
5:00 212.6 F 0.0
10:00 219.3 F 0.0
15:00 218.0 F 0.0
20:00 215.2 F 0.0
25:00 215.3 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 26:27
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:04:36
CYCLE COMPLETE

01/17/2022 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____