
Midmark M11 - v1.0.5
Total Cycles: 4443

Name: RITTER2
Sterilizer ID: V2313228
03/21/2025 02:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.4 F 0.0
2:00 128.5 F 1.3
4:00 133.0 F 3.8
6:00 169.8 F 7.4
8:00 223.6 F 11.2
10:00 223.1 F 4.6
12:00 230.0 F 9.0
14:00 239.1 F 13.4
16:00 235.8 F 10.4
18:00 236.2 F 8.7
20:00 247.7 F 14.7
22:00 258.5 F 20.9
24:00 268.8 F 27.5
26:00 271.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.9
1:00 271.9 F 30.6
2:00 273.2 F 30.7
3:00 271.5 F 30.4
4:00 272.5 F 30.9

Min 270.9 F 29.8
Max 273.6 F 31.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 226.7 F 0.0
10:00 226.3 F 0.0
15:00 229.5 F 0.0
20:00 223.0 F 0.0
25:00 225.3 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:09
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:05:07
CYCLE COMPLETE

03/21/2025 03:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____