
Midmark M11 - v1.0.0
Total Cycles: 3264

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2023 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.5 F 0.0
2:00 83.9 F 0.0
4:00 126.8 F 0.0
6:00 118.0 F 0.0
8:00 149.2 F 0.8
10:00 176.4 F 3.9
12:00 193.4 F 7.0
14:00 207.8 F 10.6
16:00 220.9 F 6.4
18:00 227.7 F 7.5
20:00 237.7 F 11.4
22:00 232.8 F 7.3
24:00 245.4 F 12.9
26:00 242.1 F 11.0
28:00 254.1 F 17.0
30:00 264.6 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.8
1:00 272.1 F 28.1
2:00 272.2 F 28.5
3:00 271.5 F 27.7
4:00 272.5 F 28.5

Min 271.2 F 27.6
Max 272.6 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 223.2 F 0.0
10:00 230.1 F 0.0
15:00 229.9 F 0.0
20:00 224.6 F 0.0
25:00 219.5 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:59
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:10:02
CYCLE COMPLETE

05/23/2023 08:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____