

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4299

Name: RITTER 1  
Sterilizer ID: V2178198  
02/13/2024 03:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.2 F 0.0  
2:00 115.0 F 1.1  
4:00 137.0 F 3.4  
6:00 173.3 F 7.3  
8:00 192.7 F 11.8  
10:00 205.1 F 3.6  
12:00 219.3 F 7.7  
14:00 230.9 F 12.1  
16:00 230.9 F 8.0  
18:00 243.2 F 13.3  
20:00 241.6 F 11.4  
22:00 253.8 F 17.5  
24:00 264.8 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.5 F 29.2  
2:00 271.5 F 28.5  
3:00 272.6 F 29.3  
4:00 271.9 F 28.7

Min 271.4 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.5 F 0.0  
5:00 216.1 F 0.0  
10:00 227.2 F 0.0  
15:00 227.0 F 0.0  
20:00 229.9 F 0.0  
25:00 230.6 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 25:53  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:04:19  
CYCLE COMPLETE

02/13/2024 04:12 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_