
Midmark M11 - v1.0.5
Total Cycles: 3209

Name: RITTER2
Sterilizer ID: V2313228
12/12/2023 04:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 84.8 F 0.0
2:00 85.7 F 0.1
4:00 98.5 F 0.8
6:00 136.1 F 3.2
8:00 166.1 F 6.7
10:00 191.3 F 10.4
12:00 209.7 F 4.7
14:00 219.6 F 7.8
16:00 230.9 F 12.3
18:00 232.1 F 8.8
20:00 240.1 F 11.0
22:00 244.1 F 12.9
24:00 256.2 F 19.3
26:00 267.2 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 273.0 F 29.9
2:00 271.7 F 29.1
3:00 272.1 F 29.8
4:00 272.6 F 29.6

Min 271.2 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 202.2 F 0.0
10:00 187.5 F 0.0
15:00 184.1 F 0.0
20:00 183.1 F 0.0
25:00 179.4 F 0.0
30:00 179.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:05:40
CYCLE COMPLETE

12/12/2023 05:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____