
Midmark M11 - v1.0.0
Total Cycles: 4176

Name: RITTER 1
Sterilizer ID: V2178198
01/03/2024 02:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.1 F 0.0
2:00 147.0 F 2.3
4:00 161.3 F 5.6
6:00 199.7 F 9.5
8:00 214.2 F 6.5
10:00 222.9 F 7.4
12:00 231.4 F 12.2
14:00 234.0 F 8.6
16:00 241.7 F 11.5
18:00 244.4 F 12.6
20:00 255.9 F 18.6
22:00 266.3 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.5 F 29.1
2:00 271.2 F 28.5
3:00 272.5 F 29.2
4:00 271.2 F 28.5

Min 271.0 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 239.5 F 0.0
10:00 230.0 F 0.0
15:00 228.1 F 0.0
20:00 225.0 F 0.0
25:00 224.5 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 23:41
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

01/03/2024 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____