
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005591
11:28 24/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:05
STERILIZING 3:03
136.0 C 220kPa 3:03
136.4 C 224kPa 3:33
136.4 C 224kPa 4:03
136.4 C 222kPa 4:33
136.4 C 225kPa 5:03
136.3 C 222kPa 5:33
136.0 C 221kPa 6:03
136.4 C 220kPa 6:33
136.4 C 220kPa 6:33
VENTING 6:33
AIR DRYING 7:22
CYCLE COMPLETE 37:22

Digital Signature #
A2671DF275C5770E
