
Midmark M11 - v1.0.0
Total Cycles: 2090

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2022 11:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.3 F 0.0
2:00 169.2 F 3.2
4:00 186.8 F 7.4
6:00 210.6 F 12.7
8:00 225.9 F 6.8
10:00 235.2 F 12.0
12:00 235.0 F 8.9
14:00 240.3 F 10.0
16:00 246.8 F 13.7
18:00 259.3 F 20.4
20:00 270.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.6 F 28.9
2:00 271.3 F 28.3
3:00 272.6 F 29.1
4:00 271.6 F 28.4

Min 271.2 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 239.2 F 0.0
10:00 226.5 F 0.0
15:00 227.0 F 0.0
20:00 223.5 F 0.0
25:00 217.5 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:41
CYCLE COMPLETE

07/05/2022 12:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____