

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002327  
9:02 26/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:37  
STERILIZING 3:39  
136.4 C 222kPa 3:39  
136.1 C 222kPa 4:09  
136.4 C 221kPa 4:39  
136.0 C 221kPa 5:09  
136.1 C 221kPa 5:39  
136.3 C 223kPa 6:09  
136.2 C 224kPa 6:39  
136.3 C 223kPa 7:09  
136.3 C 223kPa 7:09  
VENTING 7:09  
AIR DRYING 7:59  
CYCLE COMPLETE 37:59  
-----

Digital Signature #  
A646D52AEE871A90  
-----