

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2285

Name: RITTER2  
Sterilizer ID: V2313228  
04/21/2023 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.5 F 0.0  
2:00 169.4 F 3.3  
4:00 188.8 F 7.7  
6:00 209.0 F 13.2  
8:00 224.5 F 8.4  
10:00 235.3 F 11.6  
12:00 239.6 F 11.7  
14:00 239.4 F 10.5  
16:00 252.9 F 17.3  
18:00 264.7 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.5 F 29.5  
2:00 271.3 F 29.1  
3:00 272.5 F 29.5  
4:00 271.3 F 28.9  
  
Min 271.2 F 28.9  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.1 F 0.0  
5:00 216.2 F 0.0  
10:00 223.3 F 0.0  
15:00 228.8 F 0.0  
20:00 227.8 F 0.0  
25:00 228.1 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 19:51  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

04/21/2023 03:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_