
Midmark M11 - v1.0.0
Total Cycles: 3633

Name: RITTER 1
Sterilizer ID: V2178198
08/10/2023 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.7 F 0.0
2:00 174.1 F 1.4
4:00 191.1 F 5.4
6:00 212.2 F 11.0
8:00 220.3 F 5.5
10:00 234.0 F 10.8
12:00 234.0 F 7.6
14:00 243.3 F 12.0
16:00 244.8 F 12.8
18:00 257.4 F 19.4
20:00 268.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.3
3:00 272.2 F 29.0
4:00 272.4 F 28.8

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 220.3 F 0.0
10:00 238.9 F 0.0
15:00 229.3 F 0.0
20:00 223.0 F 0.0
25:00 223.1 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:11
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

08/10/2023 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____