

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000260  
13:47 30/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:25  
PRESSURIZING 3:29  
STERILIZING 4:59  
136.0 C 229kPa 4:59  
136.1 C 231kPa 5:29  
136.4 C 233kPa 5:59  
136.4 C 233kPa 6:29  
136.3 C 233kPa 6:59  
136.4 C 233kPa 7:29  
136.4 C 233kPa 7:59  
136.4 C 231kPa 8:29  
136.4 C 231kPa 8:29  
VENTING 8:30  
AIR DRYING 9:36  
CYCLE COMPLETE 24:36  
-----

Digital Signature #  
105EB2452C2441E9  
-----