

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 003372  
8:13 04/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:07  
STERILIZING 4:13  
136.5 C 221kPa 4:13  
136.4 C 223kPa 4:43  
136.4 C 223kPa 5:13  
136.4 C 219kPa 5:43  
136.3 C 222kPa 6:13  
136.3 C 223kPa 6:43  
136.4 C 223kPa 7:13  
136.0 C 221kPa 7:43  
136.0 C 221kPa 7:43  
VENTING 7:43  
AIR DRYING 8:35  
CYCLE COMPLETE 8:38

-----  
Digital Signature #  
BF622265CA395C25  
-----