
Midmark M11 - v1.0.5
Total Cycles: 453

Name: RITTER2
Sterilizer ID: V2313228
10/14/2021 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.3 F 0.0
2:00 186.7 F 4.2
4:00 205.3 F 10.2
6:00 219.4 F 6.2
8:00 236.3 F 12.6
10:00 238.8 F 11.2
12:00 239.7 F 10.8
14:00 254.3 F 18.2
16:00 266.7 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.5 F 29.5
2:00 272.6 F 30.0
3:00 272.0 F 29.1
4:00 272.9 F 30.0

Min 271.6 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 209.4 F 0.0
10:00 217.7 F 0.0
15:00 222.2 F 0.0
20:00 225.7 F 0.0
25:00 223.7 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 2:58
HEATING: 17:22
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:27
CYCLE COMPLETE

10/14/2021 01:49 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____