
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002267
9:34 15/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:45
STERILIZING 3:49
136.4 C 225kPa 3:49
136.0 C 222kPa 4:19
136.1 C 222kPa 4:49
136.3 C 223kPa 5:19
136.3 C 225kPa 5:49
136.4 C 226kPa 6:19
136.4 C 226kPa 6:49
136.3 C 225kPa 7:19
136.3 C 225kPa 7:19
VENTING 7:19
AIR DRYING 8:08
CYCLE COMPLETE 23:20

Digital Signature #
EEE395A121215D99
