

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2299

Name: RITTER 1  
Sterilizer ID: V2178198  
09/07/2022 02:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
2:00 177.6 F 0.6  
4:00 195.2 F 2.6  
6:00 212.2 F 7.6  
8:00 226.6 F 12.4  
10:00 230.6 F 8.0  
12:00 242.1 F 13.4  
14:00 241.3 F 11.3  
16:00 239.2 F 9.9  
18:00 253.2 F 16.8  
20:00 264.7 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.2  
1:00 272.4 F 28.9  
2:00 272.2 F 28.4  
3:00 272.0 F 28.6  
4:00 272.5 F 28.8  
  
Min 271.5 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 214.8 F 0.0  
10:00 227.6 F 0.0  
15:00 232.1 F 0.0  
20:00 229.1 F 0.0  
25:00 227.0 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 21:45  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:00:04  
CYCLE COMPLETE

09/07/2022 03:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_