
Midmark M11 - v1.0.5
Total Cycles: 1220

Name: RITTER2
Sterilizer ID: V2313228
06/24/2022 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.0 F 0.0
2:00 145.6 F 2.3
4:00 163.0 F 5.7
6:00 186.7 F 10.0
8:00 214.3 F 7.4
10:00 223.2 F 8.9
12:00 234.4 F 11.7
14:00 236.8 F 10.9
16:00 236.1 F 9.1
18:00 249.6 F 15.6
20:00 261.0 F 22.1
22:00 271.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 29.7
2:00 271.3 F 29.5
3:00 272.1 F 29.6
4:00 271.3 F 29.3

Min 271.2 F 29.1
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 211.1 F 0.0
10:00 229.9 F 0.0
15:00 226.3 F 0.0
20:00 227.8 F 0.0
25:00 224.0 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

06/24/2022 01:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____