
Midmark M11 - v1.0.0
Total Cycles: 3941

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2023 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 161.4 F 2.2
4:00 172.8 F 5.3
6:00 202.4 F 9.4
8:00 221.2 F 7.1
10:00 228.1 F 8.0
12:00 238.9 F 12.7
14:00 236.1 F 9.2
16:00 239.6 F 9.8
18:00 246.3 F 13.2
20:00 258.2 F 19.5
22:00 268.7 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.5 F 28.8
2:00 271.9 F 28.2
3:00 271.2 F 28.2
4:00 272.6 F 28.9

Min 271.1 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 238.3 F 0.0
10:00 223.8 F 0.0
15:00 231.6 F 0.0
20:00 230.3 F 0.0
25:00 229.3 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 23:14
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

10/24/2023 10:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____