
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000014
11:39 08/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:16
PRESSURIZING 1:23
STERILIZING 2:26
136.0 C 220kPa 2:26
136.2 C 222kPa 2:56
136.2 C 223kPa 3:26
136.3 C 222kPa 3:56
136.1 C 223kPa 4:26
136.3 C 223kPa 4:56
136.2 C 224kPa 5:26
136.3 C 220kPa 5:56
136.3 C 220kPa 5:56
VENTING 5:56
AIR DRYING 6:46
CYCLE COMPLETE 10:36

Digital Signature #
061E369A7ACA0D52
