
Midmark M11 - v1.0.0
Total Cycles: 1107

Name: RITTER 1
Sterilizer ID: V2178198
08/23/2021 07:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 79.9 F 0.0
4:00 89.6 F 0.1
6:00 136.4 F 0.1
8:00 165.4 F 0.2
10:00 192.2 F 0.3
12:00 209.8 F 3.6
14:00 222.5 F 6.5
16:00 231.3 F 9.8
18:00 239.0 F 12.1
20:00 237.3 F 9.7
22:00 238.0 F 9.2
24:00 246.2 F 13.2
26:00 243.8 F 11.9
28:00 257.0 F 18.8
30:00 268.4 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.5 F 28.4
2:00 272.0 F 28.4
3:00 272.3 F 28.3
4:00 272.1 F 28.5

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 228.2 F 0.0
10:00 194.9 F 0.0
15:00 177.5 F 0.0
20:00 215.2 F 0.0
25:00 207.1 F 0.0
30:00 200.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 31:12
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:09:15
CYCLE COMPLETE

08/23/2021 08:17 AM
ACCEPTED: X
REJECTED:
Operator: U02