
Midmark M11 - v1.0.0
Total Cycles: 2800

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2023 03:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 182.1 F 3.5
4:00 195.7 F 7.9
6:00 219.0 F 9.8
8:00 226.5 F 8.3
10:00 238.8 F 12.2
12:00 239.7 F 11.2
14:00 238.4 F 9.8
16:00 252.8 F 16.9
18:00 264.7 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 29.0
2:00 271.3 F 28.4
3:00 272.3 F 29.1
4:00 271.5 F 28.4

Min 271.3 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 232.5 F 0.0
10:00 233.7 F 0.0
15:00 220.8 F 0.0
20:00 217.6 F 0.0
25:00 218.5 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

01/19/2023 04:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____