
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003260
14:03 02/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:46
STERILIZING 3:44
136.0 C 220kPa 3:44
136.4 C 222kPa 4:14
136.4 C 225kPa 4:44
136.4 C 224kPa 5:14
136.4 C 222kPa 5:44
136.4 C 224kPa 6:14
136.4 C 225kPa 6:44
136.3 C 221kPa 7:14
136.3 C 221kPa 7:14
VENTING 7:14
AIR DRYING 8:08
CYCLE COMPLETE 8:12

Digital Signature #
CD1D04EF9E917AC9
