
Midmark M11 - v1.0.0
Total Cycles: 3747

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2023 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 173.3 F 0.0
4:00 177.2 F 2.2
6:00 198.0 F 6.1
8:00 213.4 F 10.7
10:00 222.2 F 6.3
12:00 234.4 F 11.1
14:00 232.9 F 8.1
16:00 241.8 F 11.5
18:00 245.2 F 12.8
20:00 257.2 F 19.2
22:00 268.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.0
2:00 271.7 F 28.3
3:00 272.0 F 28.9
4:00 272.2 F 28.8

Min 271.2 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 204.8 F 0.0
10:00 211.7 F 0.0
15:00 208.4 F 0.0
20:00 213.7 F 0.0
25:00 221.5 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:37
CYCLE COMPLETE

09/11/2023 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____