

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005814  
11:05 22/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:35  
PRESSURIZING 2:38  
STERILIZING 3:40  
136.5 C 220kPa 3:40  
136.4 C 224kPa 4:10  
136.4 C 221kPa 4:40  
136.4 C 223kPa 5:10  
136.4 C 224kPa 5:40  
136.4 C 225kPa 6:10  
136.4 C 225kPa 6:40  
136.4 C 224kPa 7:10  
136.4 C 224kPa 7:10  
VENTING 7:10  
AIR DRYING 7:58  
CYCLE COMPLETE 8:04

-----  
Digital Signature #  
F0157C8309E09E06  
-----