

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002408  
9:01 03/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:52  
PRESSURIZING 3:00  
STERILIZING 4:10  
136.0 C 220kPa 4:10  
136.4 C 225kPa 4:40  
136.4 C 225kPa 5:10  
136.2 C 224kPa 5:40  
136.4 C 224kPa 6:10  
136.3 C 223kPa 6:40  
136.4 C 224kPa 7:10  
136.3 C 222kPa 7:40  
136.3 C 222kPa 7:40  
VENTING 7:40  
AIR DRYING 8:31  
CYCLE COMPLETE 38:31  
-----

Digital Signature #  
5B2FCEC74EF75753  
-----