

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3593

Name: RITTER2  
Sterilizer ID: V2313228  
05/06/2024 12:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.6 F 0.0  
2:00 106.6 F 0.9  
4:00 128.7 F 3.2  
6:00 158.3 F 6.4  
8:00 185.8 F 10.4  
10:00 205.0 F 4.8  
12:00 213.3 F 7.4  
14:00 230.1 F 11.9  
16:00 231.1 F 8.4  
18:00 241.6 F 12.5  
20:00 242.1 F 12.1  
22:00 254.0 F 18.1  
24:00 264.8 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 272.5 F 30.1  
2:00 272.3 F 29.5  
3:00 271.2 F 29.1  
4:00 271.9 F 29.9

Min 271.1 F 29.0  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.0 F 0.0  
5:00 223.1 F 0.0  
10:00 194.2 F 0.0  
15:00 189.8 F 0.0  
20:00 193.9 F 0.0  
25:00 213.9 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 25:59  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:04:14  
CYCLE COMPLETE

05/06/2024 01:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_