

-----  
Midmark M11 - v1.0.5  
Total Cycles: 522

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2021 12:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.2 F 0.0  
2:00 153.0 F 2.4  
4:00 164.4 F 6.3  
6:00 192.2 F 10.9  
8:00 207.8 F 4.4  
10:00 227.7 F 10.5  
12:00 229.8 F 8.0  
14:00 239.1 F 10.9  
16:00 246.1 F 13.8  
18:00 258.5 F 20.7  
20:00 269.8 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.5 F 29.7  
2:00 271.6 F 29.4  
3:00 272.4 F 29.6  
4:00 271.5 F 29.3

Min 271.5 F 29.0  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 223.1 F 0.0  
10:00 225.1 F 0.0  
15:00 222.2 F 0.0  
20:00 208.6 F 0.0  
25:00 181.8 F 0.0  
30:00 171.0 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 20:56  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:25  
CYCLE COMPLETE

11/08/2021 01:05 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01