
Midmark M11 - v1.0.5
Total Cycles: 3435

Name: RITTER2
Sterilizer ID: V2313228
03/06/2024 12:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.4 F 0.0
2:00 99.6 F 0.0
4:00 113.1 F 0.6
6:00 140.4 F 2.9
8:00 164.6 F 6.3
10:00 185.9 F 9.9
12:00 203.4 F 9.5
14:00 215.2 F 7.1
16:00 226.7 F 11.5
18:00 231.4 F 8.2
20:00 241.7 F 12.5
22:00 242.9 F 12.4
24:00 254.5 F 18.4
26:00 265.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.7 F 30.0
2:00 271.7 F 29.3
3:00 271.4 F 29.7
4:00 272.7 F 29.8

Min 271.1 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 215.8 F 0.0
10:00 213.9 F 0.0
15:00 212.3 F 0.0
20:00 202.3 F 0.0
25:00 200.3 F 0.0
30:00 203.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 27:58
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:06:19
CYCLE COMPLETE

03/06/2024 01:56 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____