
Midmark M11 - v1.0.0
Total Cycles: 4618

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2024 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.8 F 0.0
2:00 170.8 F 3.2
4:00 182.8 F 7.2
6:00 202.9 F 11.8
8:00 216.8 F 5.5
10:00 227.9 F 10.6
12:00 229.6 F 7.0
14:00 242.1 F 13.1
16:00 241.3 F 11.5
18:00 253.8 F 17.8
20:00 264.7 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.6 F 29.5
2:00 271.8 F 29.3
3:00 272.4 F 29.4
4:00 271.8 F 29.3

Min 271.5 F 29.0
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 230.4 F 0.0
10:00 234.4 F 0.0
15:00 229.2 F 0.0
20:00 227.5 F 0.0
25:00 228.7 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:54
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

06/07/2024 12:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____