

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4523

Name: RITTER 1  
Sterilizer ID: V2178198  
05/03/2024 12:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.8 F 0.0  
2:00 163.9 F 3.0  
4:00 181.1 F 6.9  
6:00 206.8 F 11.8  
8:00 222.3 F 5.8  
10:00 234.4 F 10.7  
12:00 232.7 F 7.1  
14:00 246.1 F 13.5  
16:00 244.3 F 12.1  
18:00 256.3 F 18.6  
20:00 267.6 F 25.4  
22:00 271.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 271.8 F 28.5  
2:00 272.1 F 28.7  
3:00 272.2 F 28.9  
4:00 272.1 F 29.2

Min 271.3 F 28.5  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 196.4 F 0.0  
10:00 194.0 F 0.0  
15:00 190.8 F 0.0  
20:00 187.3 F 0.0  
25:00 186.1 F 0.0  
30:00 185.4 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 22:04  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:01:17  
CYCLE COMPLETE

05/03/2024 01:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_