
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004986
15:53 23/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:21
STERILIZING 3:26
136.5 C 220kPa 3:26
135.9 C 222kPa 3:56
136.2 C 221kPa 4:26
136.3 C 225kPa 4:56
136.5 C 225kPa 5:26
136.2 C 222kPa 5:56
136.1 C 222kPa 6:26
136.1 C 218kPa 6:56
136.1 C 218kPa 6:56
VENTING 6:57
AIR DRYING 7:48
CYCLE COMPLETE 7:55

Digital Signature #
566D7F7AB33E8D8F
