
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003884
9:06 27/06/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:12
STERILIZING 4:16
136.0 C 220kPa 4:16
136.2 C 222kPa 4:46
136.4 C 222kPa 5:16
136.2 C 218kPa 5:46
136.4 C 223kPa 6:16
136.2 C 223kPa 6:46
136.2 C 222kPa 7:16
136.4 C 222kPa 7:46
136.4 C 222kPa 7:46
VENTING 7:46
AIR DRYING 8:35
CYCLE COMPLETE 9:45

Digital Signature #
C930405FC854AB1E
