

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1767

Name: RITTER2  
Sterilizer ID: V2313228  
11/23/2022 11:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.7 F 0.0  
2:00 188.1 F 2.7  
4:00 200.7 F 8.0  
6:00 220.0 F 11.1  
8:00 228.9 F 10.0  
10:00 233.3 F 8.1  
12:00 241.1 F 11.0  
14:00 247.2 F 14.2  
16:00 260.0 F 21.3  
18:00 271.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.5 F 29.7  
2:00 272.1 F 29.1  
3:00 271.4 F 29.0  
4:00 272.5 F 29.8  
  
Min 271.3 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.0 F 0.0  
5:00 239.7 F 0.0  
10:00 227.1 F 0.0  
15:00 230.5 F 0.0  
20:00 220.5 F 0.0  
25:00 218.6 F 0.0  
30:00 217.7 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 18:43  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:50  
CYCLE COMPLETE

11/23/2022 12:00 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_