

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001141  
11:49 23/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 2:00  
STERILIZING 3:09  
136.4 C 223kPa 3:09  
136.4 C 225kPa 3:39  
136.4 C 224kPa 4:09  
136.4 C 224kPa 4:39  
136.4 C 226kPa 5:09  
136.3 C 226kPa 5:39  
136.2 C 224kPa 6:09  
136.4 C 226kPa 6:39  
136.4 C 226kPa 6:39  
VENTING 6:39  
AIR DRYING 7:31  
CYCLE COMPLETE 21:43

-----  
Digital Signature #  
DDA1340FA229CDB5  
-----