
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002593
9:10 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:19
STERILIZING 4:22
136.0 C 221kPa 4:22
136.4 C 225kPa 4:52
136.4 C 225kPa 5:22
136.3 C 224kPa 5:52
136.1 C 222kPa 6:22
136.4 C 221kPa 6:52
136.4 C 222kPa 7:22
136.4 C 224kPa 7:52
136.4 C 224kPa 7:52
VENTING 7:52
AIR DRYING 8:48
CYCLE COMPLETE 13:04

Digital Signature #
DAF3F3298F41564E
