
Midmark M11 - v1.0.5
Total Cycles: 2459

Name: RITTER2
Sterilizer ID: V2313228
06/02/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.5 F 0.0
2:00 160.2 F 2.3
4:00 166.7 F 5.7
6:00 190.4 F 9.7
8:00 214.5 F 8.6
10:00 223.7 F 8.4
12:00 235.5 F 13.3
14:00 236.3 F 10.4
16:00 236.1 F 8.7
18:00 248.6 F 15.1
20:00 259.7 F 21.4
22:00 270.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 29.8
2:00 271.0 F 29.1
3:00 272.9 F 30.2
4:00 271.7 F 29.9

Min 270.9 F 29.1
Max 273.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 217.3 F 0.0
10:00 220.5 F 0.0
15:00 220.3 F 0.0
20:00 224.0 F 0.0
25:00 218.7 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

06/02/2023 02:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____