
Midmark M11 - v1.0.0
Total Cycles: 4617

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2024 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.4 F 0.0
2:00 150.8 F 2.1
4:00 156.9 F 5.2
6:00 183.2 F 9.0
8:00 209.9 F 13.5
10:00 217.4 F 6.8
12:00 230.5 F 11.6
14:00 230.3 F 8.0
16:00 242.9 F 12.9
18:00 242.3 F 12.0
20:00 254.5 F 18.2
22:00 265.5 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.3
2:00 271.4 F 29.0
3:00 272.4 F 29.3
4:00 271.3 F 28.8

Min 271.3 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 214.2 F 0.0
10:00 228.3 F 0.0
15:00 230.1 F 0.0
20:00 226.8 F 0.0
25:00 230.6 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:46
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

06/07/2024 11:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____