

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5393

Name: RITTER 1  
Sterilizer ID: V2178198  
03/06/2025 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.0 F 0.0  
2:00 75.1 F 0.3  
4:00 90.7 F 1.7  
6:00 123.3 F 4.1  
8:00 149.9 F 6.8  
10:00 170.0 F 10.1  
12:00 192.3 F 12.4  
14:00 203.5 F 4.4  
16:00 215.2 F 7.8  
18:00 226.6 F 11.5  
20:00 225.9 F 6.3  
22:00 237.7 F 11.2  
24:00 233.9 F 8.4  
26:00 247.3 F 14.5  
28:00 258.3 F 20.5  
30:00 268.6 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.8 F 29.5  
2:00 271.7 F 28.9  
3:00 272.8 F 29.6  
4:00 272.0 F 28.9

Min 271.5 F 28.8  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
5:00 189.6 F 0.0  
10:00 188.6 F 0.0  
15:00 198.3 F 0.0  
20:00 205.5 F 0.0  
25:00 209.4 F 0.0  
30:00 212.7 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 31:14  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:09:37  
CYCLE COMPLETE

03/06/2025 10:47 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_