
Midmark M11 - v1.0.5
Total Cycles: 3144

Name: RITTER2
Sterilizer ID: V2313228
11/20/2023 12:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.8 F 0.0
2:00 186.0 F 2.3
4:00 190.5 F 6.7
6:00 207.3 F 11.2
8:00 220.4 F 5.7
10:00 231.2 F 10.7
12:00 232.6 F 7.7
14:00 242.1 F 11.8
16:00 245.4 F 13.3
18:00 257.4 F 19.8
20:00 268.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.6
2:00 271.4 F 28.8
3:00 272.5 F 29.7
4:00 272.1 F 29.2

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 225.6 F 0.0
10:00 228.0 F 0.0
15:00 228.7 F 0.0
20:00 225.6 F 0.0
25:00 222.4 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

11/20/2023 01:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____