

-----  
Midmark M11 - v1.0.5  
Total Cycles: 320

Name: RITTER2  
Sterilizer ID: V2313228  
08/30/2021 10:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.3 F 0.0  
2:00 156.7 F 2.9  
4:00 167.7 F 6.9  
6:00 192.6 F 11.9  
8:00 210.0 F 5.8  
10:00 225.5 F 10.8  
12:00 230.0 F 7.9  
14:00 240.2 F 11.8  
16:00 244.2 F 13.3  
18:00 256.6 F 19.9  
20:00 267.8 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.8 F 30.3  
2:00 272.6 F 30.1  
3:00 271.9 F 30.0  
4:00 273.1 F 30.2

Min 270.9 F 29.2  
Max 273.2 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 237.5 F 0.0  
10:00 215.3 F 0.0  
15:00 210.4 F 0.0  
20:00 206.0 F 0.0  
25:00 203.9 F 0.0  
30:00 210.5 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 21:21  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:38  
CYCLE COMPLETE

08/30/2021 12:06 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02