

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3398

Name: RITTER2  
Sterilizer ID: V2313228  
02/21/2024 04:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.3 F 0.0  
2:00 120.6 F 0.5  
4:00 134.9 F 2.8  
6:00 162.2 F 6.3  
8:00 185.2 F 10.1  
10:00 202.1 F 7.4  
12:00 214.9 F 7.3  
14:00 227.2 F 11.8  
16:00 232.9 F 8.5  
18:00 239.9 F 11.7  
20:00 244.4 F 12.9  
22:00 256.1 F 19.1  
24:00 266.6 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.5 F 29.9  
2:00 272.1 F 29.4  
3:00 270.9 F 28.9  
4:00 272.4 F 30.1

Min 270.9 F 28.9  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.4 F 0.0  
5:00 203.2 F 0.0  
10:00 190.1 F 0.0  
15:00 188.1 F 0.0  
20:00 187.5 F 0.0  
25:00 185.2 F 0.0  
30:00 183.6 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 25:42  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:04:31  
CYCLE COMPLETE

02/21/2024 05:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_