
Midmark M11 - v1.0.5
Total Cycles: 166

Name: RITTER2
Sterilizer ID: V2313228
07/06/2021 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.3 F 0.0
2:00 190.0 F 3.1
4:00 191.6 F 8.0
6:00 211.9 F 12.5
8:00 223.6 F 8.5
10:00 235.4 F 11.3
12:00 239.9 F 12.0
14:00 240.3 F 11.1
16:00 253.8 F 17.9
18:00 265.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.0 F 29.3
2:00 272.1 F 29.7
3:00 271.3 F 28.9
4:00 272.1 F 29.6

Min 271.1 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 229.6 F 0.0
10:00 229.0 F 0.0
15:00 208.4 F 0.0
20:00 191.0 F 0.0
25:00 187.2 F 0.0
30:00 186.9 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:41
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

07/06/2021 10:20 AM
ACCEPTED: X
REJECTED:
Operator: U02