

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002221  
12:37 05/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:10  
PRESSURIZING 3:13  
STERILIZING 4:22  
136.0 C 222kPa 4:22  
136.3 C 225kPa 4:52  
136.3 C 223kPa 5:22  
136.4 C 226kPa 5:52  
136.4 C 224kPa 6:22  
136.4 C 224kPa 6:52  
136.4 C 226kPa 7:22  
136.4 C 226kPa 7:52  
136.4 C 226kPa 7:52  
VENTING 7:52  
AIR DRYING 8:48  
CYCLE COMPLETE 24:25

-----  
Digital Signature #  
56186CB6CDD44DC1  
-----