
Midmark M11 - v1.0.5
Total Cycles: 2158

Name: RITTER2
Sterilizer ID: V2313228
03/16/2023 03:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.6 F 0.0
2:00 168.0 F 2.8
4:00 189.9 F 6.5
6:00 206.9 F 11.1
8:00 220.7 F 4.7
10:00 229.8 F 9.9
12:00 233.8 F 8.6
14:00 242.0 F 12.4
16:00 240.8 F 11.0
18:00 253.5 F 17.4
20:00 264.3 F 24.1
22:00 271.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.7 F 29.6
2:00 271.7 F 29.5
3:00 272.4 F 29.6
4:00 271.5 F 29.4

Min 271.3 F 29.1
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 230.0 F 0.0
10:00 230.8 F 0.0
15:00 216.5 F 0.0
20:00 225.1 F 0.0
25:00 228.9 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

03/16/2023 04:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____