

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3304

Name: RITTER 1  
Sterilizer ID: V2178198  
05/30/2023 11:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.9 F 0.0  
2:00 142.0 F 0.0  
4:00 168.6 F 0.0  
6:00 191.2 F 4.3  
8:00 211.6 F 9.0  
10:00 224.9 F 10.6  
12:00 231.5 F 8.9  
14:00 242.0 F 11.3  
16:00 242.0 F 11.7  
18:00 240.1 F 10.3  
20:00 253.9 F 17.2  
22:00 265.7 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.5 F 28.6  
2:00 271.8 F 28.5  
3:00 272.4 F 28.6  
4:00 271.6 F 28.3

Min 271.5 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.4 F 0.0  
5:00 215.7 F 0.0  
10:00 229.4 F 0.0  
15:00 230.5 F 0.0  
20:00 229.4 F 0.0  
25:00 229.1 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 23:41  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:01:46  
CYCLE COMPLETE

05/30/2023 12:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_