

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1833

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2022 02:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.4 F 0.0  
2:00 161.1 F 2.7  
4:00 170.8 F 6.7  
6:00 197.4 F 12.0  
8:00 217.0 F 6.5  
10:00 231.8 F 11.8  
12:00 233.8 F 9.2  
14:00 240.0 F 9.7  
16:00 247.4 F 14.6  
18:00 260.0 F 21.6  
20:00 271.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.6 F 29.8  
2:00 272.0 F 29.2  
3:00 271.5 F 29.5  
4:00 272.7 F 29.8

Min 271.4 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 211.5 F 0.0  
10:00 227.1 F 0.0  
15:00 234.1 F 0.0  
20:00 229.1 F 0.0  
25:00 226.3 F 0.0  
30:00 218.1 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 20:41  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:29  
CYCLE COMPLETE

12/13/2022 03:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_